



**Doctor Vida**

PRODUCTS

Powered by **STAB VIDA**

Your Easy DNA Laboratory

**Ref. 133001017**



OVERVIEW .....	3
KIT CONTENTS.....	3
SHIPPING AND STORAGE.....	3
PRODUCT USE LIMITATIONS .....	3
SAFETY INFORMATION.....	3
EQUIPMENTS AND REAGENTS NOT PROVIDED WITH THE PRODUCT .....	4
PROTOCOL .....	4
MANUFACTURER INFORMATION.....	4

The Doctor Vida Lysis Buffer is a solution containing detergents designed for effective cell disruption and protection of DNA molecules from degradation. This product is only compatible with Doctor Vida Customized LAMP Master Mix (ref. 133001010) and Universal LAMP Master Mix (ref. 133001018).

Doctor Vida Lysis Buffer is suitable for cell disruption of several types of samples, such as, environment, plants, humans, and animals.

Doctor Vida Lysis Buffer (ref. 133001017).

The product is shipped and stored at 2-8°C.

This product is only compatible with Doctor Vida Customized LAMP Master Mix (ref. 133001010) and Universal LAMP Master Mix (ref. 133001018).

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the safety data sheet (SDSs) available online in convenient and compact PDF format at <https://doctorvida.store/>. STAB VIDA recommends to have available the contacts of medical emergency and poison center for all staff members.

- Pipets and pipet tips (aerosol resistant)
- Laboratory consumables
- Doctor Vida Customized LAMP Master Mix (ref. 133001010) or Universal LAMP Master Mix (ref. 133001018).

1. Dispense 200µl of lysis buffer in the 1.5ml microcentrifuge tubes.

Important note 1: The volume of lysis buffer used to resuspend the sample should be optimized based on the sample type. Higher volumes of lysis buffer might be required.

2. Add the sample (raw material or 1-5 uL of extracted DNA) to the Doctor Vida lysis buffer. Vortex in order to homogenize. Use this solution as “Lysate”.

Important note 2: A minimum of 10 ng of DNA sample is recommended to be added to the lysis buffer.

Important note 3: Solid samples, such as plant leaves, may need to be crushed in water before being added to the lysis buffer.

Important note 4: The amount of sample to be added to the lysis buffer should be optimized.

3. Incubate the lysate (sample in lysis buffer) 10 - 30 minutes at room temperature before proceeding. Lysate can be immediately used after incubation or can be stored for maximum 24h in the fridge.

: STAB VIDA- Research and Services in Biological Sciences, Lda.  
Madan Parque, Rua dos Inventores, Sala 2.18, 2825-182 Caparica, Portugal.

[www.stabvida.com](http://www.stabvida.com)



**+351 938 437 766**

In case of any problem, please contact us by email [drvida@stabvida.com](mailto:drvida@stabvida.com) or phone 00351 938 437 766 (Call to national mobile network)  
Monday to Friday from 8:30 am to 5:30 pm. (GMT time)